

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



0325

Accredited to
ISO/IEC 17025:2017

Precision Technologies International Ltd t/a Precision Technologies

Issue No: 019 Issue date: 25 March 2020

Mariner
Lichfield Road Industrial Estate
Tamworth
Staffordshire
B79 7UL

Contact: Mr L Hyland
Tel: +44 (0)1827 315 050
Fax: +44 (0)1827 317 921
E-Mail: enquiries@ptilttd.co.uk
Website: http://precisiontechnologies.co.uk

Calibration performed at the above address only

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k=2)	Remarks
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED			
GEARS SPUR/HELICAL EXTERNAL			NOTES
External (Tip) Diameter	10 to 100 100 to 200 200 to 300	1.2] Even 2.0] Odd 3.8 No of 5.0 No of 5.0] Teeth 6.5] Teeth	Horizontal measuring machine and reference setting standards.
Bore Diameter	12 to 100 100 to 200 200 to 250	1.2 3.8 5.0	
Profile total deviation (F_a)	0.6 module to 3 module 3 module to 6 module	1.5 2.5	CNC gear measuring machine.
Helix (lead) total deviation (F_β) /Helix Angle	0 to 50 Facewidth 0° to 15° 15° to 30° Helix Angle 30° to 45° 50 to 100 Facewidth 0° to 15° 15° to 30° Helix Angle 30° to 45°	1.5 2.0 2.5 2.0 2.5 3.0	
Cumulative pitch deviation (F_p)	Min approx 1 module to 300 diameter	2.0	
Single pitch deviation (f_p)	Min approx 1 module to 300 diameter	2.0	
Adjacent pitch difference (f_u)	Min approx 1 module to 300 diameter	2.0	
Radial runout of tooth space deviation (F_r)	10 to 300	1.5	
Dimension over pins or rollers Spur gears (Mdr)	10 to 100 diameter 100 to 200 diameter 200 to 300 diameter	5.0 7.5 10	Horizontal measuring machine and reference setting standards.



0325
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees , Chertsey Lane, Staines-upon-Thames , TW18 3HR, UK

Precision Technologies International Ltd t/a Precision Technologies
Issue No: 019 Issue date: 25 March 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k=2)	Remarks
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED			
GEARS SPUR/HELICAL EXTERNAL (cont'd)			
Dimension over pins or rollers Helical gears (Mdr)	10 to 100 diameter	6.0	Horizontal measuring machine and reference setting standards and calibrated pins.
	100 to 200 diameter	8.5	
	200 to 300 diameter	12	
INTERNAL			
As for external gears except for:			
Internal (Tip) Diameter	22 to 100	1.5 } Even 2.5 } Odd	CNC gear measuring machine.
	100 to 200	4.0 } No of 5.5 } No of	
	200 to 250	5.0 } Teeth 7.0 } Teeth	
Dimension under Rollers or Pins (Mdr)	20 to 100	5.0 } Spur 6.0 } Helical	Horizontal measuring machine and reference setting standards.
	100 to 200	7.5 } Gears 8.5 } Gears	
	200 to 250	10 } 12 }	
SPLINE GAUGES, INVOLUTE SPUR/HELICAL EXTERNAL/INTERNAL			
As for Gears except for:			
Helix (lead) total deviation (F _β) /Helix Angle	0 to 50 Facewidth	1.5	CNC gear measuring machine.
	0° to 15° Helix 15° to 30° Angle		
Runout Major/Minor Diameter	50 to 100 Facewidth	2.0	Horizontal measuring machine and reference setting standards.
	0° to 15° Helix 15° to 30° Angle		
SPLINE GAUGES, STRAIGHT SIDED PLUG AND RING			
As for Involute Spline Gauges			
Runout of Major/Minor Diameters	0 to 300	2.5	Horizontal measuring machine and reference setting standards.
Spline Width		1.2	
SERRATION GAUGES STRAIGHT SIDED PLUG AND RING			
As for straight sided Spline Gauges			
Angle of Serration Flank Spur Gauges Only	0 to 300	15 minutes of arc	Horizontal measuring machine and reference setting standards and calibrated pins.



0325

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees , Chertsey Lane, Staines-upon-Thames , TW18 3HR, UK

Precision Technologies International Ltd t/a Precision Technologies

Issue No: 019 Issue date: 25 March 2020

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks
END			



0325

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Precision Technologies International Ltd t/a Precision Technologies

Issue No: 019 Issue date: 25 March 2020

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are not mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %·V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %· p + (0.12·10⁻⁶· p ·10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i , where i is the instrument indication.